Arlington, TX 76012
Pocomoke City, MD 21851
Batesville, IN 47006
http://www.riccachemical.com
1-888-GO-RICCA
customerservice@riccachemical.com

## **Product Specification**

Bromocresol Green Indicator, 0.1% (w/v)

Lot Number: SAMPLE Product Number: 1205

Manufacture Date: N/A Expiration Date: N/A

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Bromocresol Green, Sodium Salt	62625-32-5	ACS

Test	Specification	Result
Appearance	Dark blue liquid	N/A
Suitability for Use	Blue (pH >5.4) - Green (pH 4.5) - Yellow (pH < 4.0)	N/A

Specification	Reference
Bromcresol Green Indicator Solution (1 g/L)	ASTM (D 1067 B)
Bromcresol green indicator solution, pH 4.5 indicator	АРНА (2320 В)
Bromcresol Green Indicator Solution	APHA (3500-Al B)
Bromcresol Green Indicator Solution	APHA (7500-Ra B)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
1205-1	4 L natural poly	24 months
1205-16	500 mL natural poly	24 months
1205-32	1 L natural poly	24 months
1205-4	120 mL natural poly	24 months
R1205000-55E	55 gal poly drum	12 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

Version: 1.3 Lot Number: SAMPLE Product Number: 1205 Page 1 of 1