

Product Specification

Crystal Violet Indicator, 1% (w/v) in Glacial Acetic Acid Gentian Violet Indicator for Nonaqueous Titrations

Lot Number: SAMPLE

Product Number: 2483

Manufacture Date: N/A

Expiration Date: N/A

Name	CAS#	Grade
Acetic Acid	64-19-7	ACS
Gentian Violet	548-62-9	ACS

Test	Specification	Result
Appearance	Dark blue to violet/purple liquid	N/A
Suitability for Use	Purple (Basic) - Blue-Green (Acidic)	N/A

Specification	Reference
Crystal Violet Indicator Solution (10 g/L)	ASTM (E 200)
Crystal Violet Indicator Reagent Solution	ACS (N/A)
Crystal Violet TS	USP (N/A)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
2483-16	500 mL amber glass	12 months
2483-4	120 mL amber glass	12 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.