

Product Specification

Gardner Color Standard #14

Lot Number: SAMPLE

Product Number: 3226.14

Manufacture Date: N/A

Expiration Date: N/A

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Ferric Chloride Hexahydrate	10025-77-1	ACS
Cobalt (II) Chloride Hexahydrate	7791-13-1	ACS
Hydrochloric Acid	7647-01-0	ACS

Test	Specification	Result
Appearance	Yellow liquid	N/A
Gardner color (ASTM D 6166)	13.7-14.3	N/A

Specification	Reference
---------------	-----------

Gardner Color Standard No. 14

ASTM (D 1544-58 T)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
3226.14-100	100 mL amber glass	12 months
3226.14-16	500 mL amber glass	12 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.