

## **Product Specification**

## Low Level Standard

Lot Number: SAMPLE

Product Number: RCELLOW0

## Manufacture Date: N/A Expiration Date: N/A

Matrix: 2% Nitric Acid / 0.05% HCl

Name		CAS#	Grade		
Water		7732-18-5	ACS/ASTM/USP/EP		
Nitric Acid		7697-37-2	Trace Metals		
Hydrochloric Acid		7647-01-0	Trace Metals		
Test		Specification	Result		
Appearance		Colorless liquid	N/A		
Analyte	Analysis (ppm)	Solute	CAS#	Grade	NIST SRM#
Aluminum (Al)	24.7-25.3 ppm	Aluminum Nitrate Nonahydrate	7784-27-2	High Purity	3101
Iron (Fe)	24.7-25.3 ppm	Iron	7439-89-6	High Purity	3126
Chromium (Cr)	0.99-1.01 ppm	Chromium Nitrate Nonahydrate	7789-02-8	High Purity	3112
Selenium (Se)	0.99-1.01 ppm	Selenium	7782-49-2	High Purity	3149
Zinc (Zn)	0.99-1.01 ppm	Zinc	7440-66-6	High Purity	3168
Copper (Cu)	0.99-1.01 ppm	Copper	7440-50-8	High Purity	3114
Nickel (Ni)	0.99-1.01 ppm	Nickel	7440-02-0	High Purity	3136
Balances are calibrate Operating Procedures.	d using NIST traceable we . Class A glassware is also	urate provided the product is kept tight ights whose verification of maintenanc calibrated and routinely rechecked per r are used in the manufacture of this pu	e and recalibration in house Standard	is documented per in-he Operating Procedures.	ouse Standard Trace metal

analyzed acids and Trace Metals Analyzed Water are used in the manufacture of this product. Triple cleaned containers are used in the manufacture of this product.

 Part Number
 Size / Package Type
 Shelf Life (Unopened Container)

	Fart Number	Size / Package Type	Shelf Life (Unopened Container)			
	RCELLOW0-500N	500 mL natural poly	12 months			
Recommended Storage: 15°C - 30°C (59°F - 86°F)						

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.