Arlington, TX 76012
Pocomoke City, MD 21851
Batesville, IN 47006
http://www.riccachemical.com
1-888-GO-RICCA
customerservice@riccachemical.com

Product Specification

Instrument Calibration Standard 1 - CLP

Lot Number: SAMPLE Product Number: RINSTCL1

Manufacture Date: N/A Expiration Date: N/A

This is a multielement solution that was prepared gravimetrically to contain the elements and concentrations stated below. The uncertainty associated with the certified values is $\pm 0.5\%$ relative. The concentrations are confirmed by ICP or ICP-MS.

Matrix: 5% Nitric Acid

Name	CAS#	Grade	
Water	7732-18-5	ACS/ASTM/USP/EP	
Nitric Acid	7697-37-2	Trace Metals	

Test	Specification	Result	
Appearance	Colorless liquid	N/A	

Analyte	Analysis (ppm)	Solute	CAS#	Grade	NIST SRM#
Calcium (Ca)	4975-5025 ppm	Calcium Carbonate	471-34-1	High Purity	3109
Sodium (Na)	4975-5025 ppm	Sodium Carbonate	497-19-8	High Purity	3152
Potassium (K)	4975-5025 ppm	Potassium Carbonate	584-08-7	High Purity	3141
Magnesium (Mg)	4975-5025 ppm	Magnesium	7439-95-4	High Purity	3131

Specification		Reference	
	Instrument Calibration Standard 1, CLP 5000 ppm each Ca, Mg,	EPA (200.7)	
	K, Na		
	Instrument Calibration Standard 1, CLP 5000 ppm each Ca, Mg,	EPA (SW-846) (6010)	
	K, Na		

This standard is guaranteed to be stable and accurate provided the product is kept tightly capped and stored under normal laboratory conditions. Balances are calibrated using NIST traceable weights whose verification of maintenance and recalibration is documented per in-house Standard Operating Procedures. Class A glassware is also calibrated and routinely rechecked per in-house Standard Operating Procedures. Trace metal analyzed acids and Trace Metals Analyzed Water are used in the manufacture of this product. Triple cleaned containers are used in the manufacture of this product.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
RINSTCL1-100	100 mL natural LDPE	12 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

Version: 1.3 Lot Number: SAMPLE Product Number: RINSTCL1 Page 1 of 2

