Arlington, TX 76012
Pocomoke City, MD 21851
Batesville, IN 47006
http://www.riccachemical.com
1-888-GO-RICCA
customerservice@riccachemical.com

Product Specification

Boron ICP Standard, 1000 ppm B in H₂O/tr NH₄OH

Lot Number: SAMPLE Product Number: PB1KW

Iridium (Ir)

Lanthanum (La)

Iron (Fe)

N/A

N/A

N/A

Manufacture Date: N/A Expiration Date: N/A

This is a single element solution that was prepared volumetrically to contain the certified value reported. The uncertainty associated with the certified value is the sum of the estimated errors due to the purity of the raw material, the volumetric preparation of the solution, and transpiration of the solution through the container wall.

The final solution concentration is confirmed by AA, ICP, or ICP-MS, and is traceable to NIST Standard Reference Material 3107. All trace level elements were determined by ICP or ICP-MS.

Name		CAS# Grae		drade		
Water		7732-18	3-5			
Boric Acid		10043-35-3				
Ammonium Hydroxide		1336-21-6				
Test		Chasi	fication	Result	NIST SRM#	_
rest					THE SIGNE	
Appearance		Color	less liquid	N/A		
Boron (B)		995-1	1005 ppm	N/A	3107	
		Trace Elements by		S I=Spectral	Interference N=Not Tested	1
Aluminum (Al)	N/A	Lead (Pb)	N/A	Strontium (Sr)	N/A	
Antimony (Sb)	N/A	Lithium (Li)	N/A	Sulfur (S)	N/A	
Arsenic (As)	N/A	Lutetium (Lu)	N/A	Tantalum (Ta)	N/A	
Barium (Ba)	N/A	Magnesium (Mg)	N/A	Tellurium (Te)	N/A	
Beryllium (Be)	N/A	Manganese (Mn)	N/A	Terbium (Tb)	N/A	
Bismuth (Bi)	N/A	Mercury (Hg)	N/A	Thallium (Tl)	N/A	
Cadmium (Cd)	N/A	Molybdenum (Mo)	N/A	Thorium (Th)	N/A	
Calcium (Ca)	N/A	Neodymium (Nd)	N/A	Thulium (Tm)	N/A	
Cerium (Ce)	N/A	Nickel (Ni)	N/A	Tin (Sn)	N/A	
Cesium (Cs)	N/A	Niobium (Nb)	N/A	Titanium (Ti)	N/A	
Chromium (Cr)	N/A	Osmium (Os)	N/A	Tungsten (W)	N/A	
Cobalt (Co)	N/A	Palladium (Pd)	N/A	Uranium (U)	N/A	
Copper (Cu)	N/A	Phosphorus (P)	N/A	Vanadium (V)	N/A	
Dysprosium (Dy)	N/A	Platinum (Pt)	N/A	Ytterbium (Yb)	N/A	
Erbium (Er)	N/A	Potassium (K)	N/A	Yttrium (Y)	N/A	
Europium (Eu)	N/A	Praseodymium (Pr)	N/A	Zinc (Zn)	N/A	
Gadolinium (Gd)	N/A	Rhenium (Re)	N/A	Zirconium (Zr)	N/A	
Gallium (Ga)	N/A	Rhodium (Rh)	N/A			
Germanium (Ge)	N/A	Rubidium (Rb)	N/A			
Gold (Au)	N/A	Ruthenium (Ru)	N/A			
Hafnium (Hf)	N/A	Samarium (Sm)	N/A			
Holmium (Ho)	N/A	Scandium (Sc)	N/A			
Indium (In)	N/A	Selenium (Se)	N/A			

Version: 1.3 Lot Number: SAMPLE Product Number: PB1KW Page 1 of 2

N/A

N/A

N/A

Silicon (Si)

Silver (Ag)

Sodium (Na)

Specification	Reference
Boron ICP, 1000 ppm in H2O	EPA (200.7)

This standard is guaranteed to be stable and accurate provided the product is kept tightly capped and stored under normal laboratory conditions. Balances are calibrated using NIST traceable weights whose verification of maintenance and recalibration is documented per in-house Standard Operating Procedures. Class A glassware is also calibrated and routinely rechecked per in-house Standard Operating Procedures. Trace metal analyzed acids and Trace Metals Analyzed Water are used in the manufacture of this product. Triple cleaned containers are used in the manufacture of this product.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
PB1KW-100	100 mL natural LDPE	18 months
PB1KW-250	250 mL natural poly	18 months
PB1KW-500	500 mL natural poly	18 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

This document is designed to comply with ISO Guide 31 "Reference Materials $^{\hbox{--}}$ Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

Version: 1.3 Lot Number: SAMPLE Product Number: PB1KW Page 2 of 2