

Certificate of Analysis

Buffer, Reference Standard, pH 10.00 ± 0.01 at 25°C (Color Coded Blue)

Lot Number: 1308J86

Product Number: 1601

Manufacture Date: AUG 11, 2023

Expiration Date: JAN 2025

The certified value for this product is confirmed in independent testing by a second qualified chemist.

The NIST traceable pH value is certified to ±0.01 at 25 °C only. All other pH values at their corresponding temperatures are accurate to ± 0.05.

| °C | 0 | 5 | 10 | 15 | 20 | 25 | 30 | 35 | 40 | 50 |
|----|-------|-------|-------|-------|-------|-------|------|------|------|------|
| pH | 10.31 | 10.23 | 10.17 | 10.11 | 10.05 | 10.00 | 9.95 | 9.91 | 9.87 | 9.81 |

| Name | CAS# | Grade |
|--------------------|-------------|-----------------|
| Water | 7732-18-5 | ACS/ASTM/USP/EP |
| Sodium Carbonate | 497-19-8 | ACS |
| Sodium Bicarbonate | 144-55-8 | ACS |
| Sodium Hydroxide | 1310-73-2 | Reagent |
| Preservative | Proprietary | |
| Blue Dye | Proprietary | |

| Test | Specification | Result |
|------------|---------------|--------------------------------|
| Appearance | Blue liquid | Passed *Not a certified value. |

| Test | Certified Value | Uncertainty | NIST SRM# |
|---------------------------------------|-----------------|-------------|-------------------------|
| pH at 25°C (Method: SQCP027, SQCP033) | 10.008 | 0.02 | 186-I-g, 186-II-g, 191d |

| Specification | Reference |
|-----------------------------|-----------------|
| Commercial Buffer Solutions | ASTM (D 1293 B) |
| Buffer C | ASTM (D 5464) |
| Buffer C | ASTM (D 5128) |

pH measurements were performed in our Arlington, TX laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution. Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

| Part Number | Size / Package Type | Shelf Life (Unopened Container) |
|-------------|---------------------|---------------------------------|
| 1601-2.5 | 10 L Cubitainer® | 18 months |
| 1601-5 | 20 L Cubitainer® | 18 months |

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Heidi J Green (08/11/2023)

Operations Manager

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This product was tested in an ISO 17025 Accredited Laboratory

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.