

# Certificate of Analysis

## EP Reference Solution BY<sub>3</sub>, Brownish-Yellow Reference Solution

**Lot Number:** 1308K72

**Product Number:** 2873

**Manufacture Date:** AUG 16, 2023

**Expiration Date:** AUG 2024

The certified value reported is the prepared value based upon the method of preparation of the material. The uncertainty in the prepared value is based upon the volumetric method of preparation.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Hydrochloric Acid	7647-01-0	ACS
Ferric Chloride Hexahydrate	10025-77-1	ACS
Cobalt (II) Chloride Hexahydrate	7791-13-1	ACS
Copper Sulfate Pentahydrate	7758-99-8	ACS

Test	Specification	Result
Appearance	Brownish-yellow liquid	Passed

Specification	Reference
Standard Solution BY <sub>3</sub> (brownish-yellow)	EP (2.2.2)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
2873-4	120 mL amber glass	12 months

**Recommended Storage:** 15°C - 30°C (59°F - 86°F)



Heidi J Green (08/16/2023)

Operations Manager

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.