

Certificate of Analysis

Turbidity Standard, 3 NTU, Polymer Bead Suspension

Lot Number: 1403H08

Product Number: R8805300

Manufacture Date: MAR 14, 2024

Expiration Date: MAR 2025

Ready to use. Do NOT agitate. Pour gently into cuvette avoiding bubbles which will interfere with Turbidity readings. There is no NIST Standard Reference Material for Turbidity. This solution is traceable to commercially prepared EPA recognized Primary Standards. The EPA primary standard for turbidity (Formazin suspension, 4000 NTU) is identical to EP Primary Opalescent Suspension and USP Formazin Stock Suspension.

| Name | CAS# | Grade |
|--------------------------|-------------|-----------------|
| Water | 7732-18-5 | ACS/ASTM/USP/EP |
| Preservative | Proprietary | |
| Polystyrene Microspheres | 9003-53-6 | Reagent |

| Test | Specification | Result |
|------------|------------------|---------|
| Appearance | Colorless liquid | Passed |
| Turbidity | 2.9-3.1 NTU | 3.0 NTU |

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

| Part Number | Size / Package Type | Shelf Life (Unopened Container) |
|---------------|---------------------|---------------------------------|
| R8805300-100C | 100 mL amber glass | 12 months |
| R8805300-1C | 1 L amber glass | 12 months |
| R8805300-250C | 250 mL amber glass | 12 months |

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Heidi J Green (03/14/2024)

Operations Manager

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.