

Certificate of Analysis

Bromate-Bromide Solution, 0.500 Normal (N/2), 0.250 Molar as Br₂

Lot Number: 1403L74

Product Number: 1172

Manufacture Date: MAR 28, 2024

Expiration Date: MAR 2026

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Potassium Bromide	7758-02-3	ACS
Potassium Bromate	7758-01-2	ACS

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	Passed	
Assay (vs. Potassium Iodate/Starch)	0.4998-0.5002 N at 20°C	0.5000 N at 20°C	136

Specification	Reference
---------------	-----------

Bromide-Bromate, Standard Solution (0.2500 M as Br₂)

ASTM (D 1159)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
1172-32	1 L amber glass	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Heidi J Green (03/28/2024)

Operations Manager

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.