

Certificate of Analysis

Potassium Thiocyanate, 0.100 Normal (N/10)

Lot Number: 1404E55

Product Number: 6580

Manufacture Date: APR 05, 2024

Expiration Date: SEP 2025

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Potassium Thiocyanate	333-20-0	ACS

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	Passed	
Assay (vs. Silver Nitrate/Volhard's Indicator)	0.0999-0.1001 N at 20°C	0.1000 N at 20°C	999

Specification	Reference
Potassium Thiocyanate, 0.1 N	TAPPI (T 624 cm-85)
Potassium Thiocyanate, 0.1 N	TAPPI (T 610 sp-97)
Potassium Thiocyanate, 0.1 Normal Volumetric Solution	ACS (N/A)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
6580-1	4 L natural poly	18 months
6580-32	1 L natural poly	18 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Heidi J Green (04/05/2024)

Operations Manager

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.