448 West Fork Dr Arlington, TX 76012 http://www.riccachemical.com 1-888-GO-RICCA

customerservice@riccachemical.com

## Certificate of Analysis

## Turbidity Standard, Primary Standard Stock Formazin Suspension, 4000 NTU

Lot Number: 1404F73 Product Number: 8830

Manufacture Date: APR 12, 2024

Expiration Date: APR 2025

Ready to use. Do NOT agitate. Pour gently into cuvette avoiding bubbles which will interfere with Turbidity readings.

There is no NIST Standard Reference Material for Turbidity. This solution is traceable to commercially prepared EPA recognized Primary Standards.

The EPA primary standard for turbidity (Formazin suspension, 4000 NTU) is identical to EP Primary Opalescent Suspension and USP Formazin Stock Suspension.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Hydrazine Sulfate	10034-93-2	ACS
Hexamethylenetetramine	100-97-0	ACS

Test	Specification	Result
Appearance	Milky white suspension	Passed
Turbidity	3800-4200 NTU	4025 NTU

Specification	Reference
Stock primary standard formazin suspension	АРНА (2130 В)
Primary Opalescent Suspension	USP (N/A)
Primary opalescent suspension	EP (2.2.1)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
8830-16	500 mL amber glass	12 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

Heidi J Green (04/12/2024)

**Operations Manager** 

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

Version: 1.3 Lot Number: 1404F73 Product Number: 8830 Page 1 of 1