

Certificate of Analysis

Silver Nitrate, 0.0100 Normal (N/100)

Lot Number: 1404G18

Product Number: 6850

Manufacture Date: AUG 31, 2023

Expiration Date: MAR 2026

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Silver Nitrate	7761-88-8	ACS

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	Passed	
Assay (vs. Potassium Chloride/Potassium Chromate)	0.00998-0.01002 N at 20°C	0.01000 N at 20°C	999

Specification	Reference
Silver Nitrate Solution, Standard (0.01 N)	ASTM (D 2384)
Silver Nitrate Solution, Standard (0.01 N)	ASTM (D 2036 A)
Silver Nitrate Solution, Standard (0.01 N)	ASTM (D 2036 B)
Silver Nitrate Solution, Standard (0.01 N)	ASTM (D 2036 C)
Silver Nitrate Solution, Standard (0.01 N)	ASTM (D 2036 D)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
6850-1	4 L black poly	24 months
6850-32	1 L amber poly	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Heidi J Green (08/31/2023)

Operations Manager

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.