

Certificate of Analysis

Sheather's Sugar Solution, Sp. Gr. 1.250, for Flotation of Parasites

Lot Number: 1404G48

Product Number: 6713

Manufacture Date: APR 12, 2024

Expiration Date: APR 2025

Name	CAS#	Grade
Sucrose	57-50-1	Food
Water	7732-18-5	ACS/ASTM/USP/EP
Formaldehyde	50-00-0	ACS
Methyl Alcohol	67-56-1	ACS

Test	Specification	Result
Appearance	Yellowish-gold liquid	Passed
Specific Gravity	1.249-1.251 at 25°C	1.251 at 25°C

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
6713-32	1 L natural poly	12 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

Heidi J Green (04/12/2024)

Operations Manager

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.