

# Certificate of Analysis

## New Methylene Blue "N" Stain, for Reticulocyte Count

**Lot Number:** 1404J51

**Product Number:** 5285

**Manufacture Date:** APR 22, 2024

**Expiration Date:** APR 2026

For In Vitro Diagnostic Use. This product is suitable for use in identification and enumeration of reticulocytes in peripheral blood. A recommended procedure is available in RICCA CHEMICAL COMPANY Technical Reference Document TRD15. This product may also be suitable for use in other methods.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Sodium Chloride	7647-14-5	ACS
New Methylene Blue N	6586-05-6	Certified
Potassium Oxalate Monohydrate	6487-48-5	ACS

Test	Specification	Result
Appearance	Dark blue liquid	Passed
UV/Visible Spectrum	To Pass Test	Passed

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
5285-16	500 mL natural poly	24 months
5285-4	120 mL natural poly	24 months

**Recommended Storage:** 15°C - 30°C (59°F - 86°F)



Heidi J Green (04/22/2024)  
 Operations Manager

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.