

# Certificate of Analysis

**Silver Nitrate, 0.141 Normal, 1 mL = 5 mg Cl<sup>-</sup>**

**Lot Number:** 1405C23

**Product Number:** 7002

**Manufacture Date:** MAY 01, 2024

**Expiration Date:** APR 2026

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Silver Nitrate	7761-88-8	ACS

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	Passed	
Assay (vs. Potassium Chloride/Potassium Chromate)	0.1409-0.1411 N at 20°C	0.1411 N at 20°C	999

Specification	Reference
---------------	-----------

Silver Nitrate Solution, Standard (equivalent to 5 mg Cl<sup>-</sup>/mL)      ASTM (D 4458)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
7002-1	4 L black poly	24 months
7002-16	500 mL amber poly	24 months
7002-32	1 L amber poly	24 months

**Recommended Storage:** 15°C - 30°C (59°F - 86°F)



Heidi J Green (05/01/2024)  
 Operations Manager

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.