

Certificate of Analysis

Methyl Red-Methylene Blue TS, Alcoholic Mixed Indicator

Lot Number: 1405C79

Product Number: 5090

Manufacture Date: MAY 02, 2024

Expiration Date: OCT 2025

Name	CAS#	Grade
Ethyl Alcohol	64-17-5	Reagent
Methyl Alcohol	67-56-1	Reagent
Methylene Blue Trihydrate	7220-79-3	Certified
Methyl Red	493-52-7	ACS

Test	Specification	Result
Appearance	Dark green liquid	Passed
Suitability for Use	Red-Violet (pH < 5.2) - Green (pH > 5.6)	Passed

Specification	Reference
Methyl Red-Methylene Blue TS	USP (N/A)
Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.	

Part Number	Size / Package Type	Shelf Life (Unopened Container)
5090-16	500 mL natural poly	18 months
5090-4	120 mL natural poly	18 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



 Heidi J Green (05/02/2024)
 Operations Manager

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.