

Certificate of Analysis

Citric Acid, 1 gram per 1 mL Water, Filtered

Lot Number: 2309J49

Product Number: 2126

Manufacture Date: SEP 19, 2023

Expiration Date: SEP 2025

Name	CAS#	Grade
Citric Acid	77-92-9	USP
Water	7732-18-5	ACS/ASTM/USP/EP

Test	Specification	Result	NIST SRM#
Appearance	Colorless to slightly yellow liquid	Passed	
Assay (vs. Sodium Hydroxide/Phenolphthalein)	49.95-50.05 % (w/w)	49.99 % (w/w)	84

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
2126-1	4 L natural poly	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Sharon Travers (09/19/2023)

Operations Manager

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.