

Certificate of Analysis

Sulfuric Acid, 0.0500 Normal (N/20)

Lot Number: 2309L15

Product Number: 8219

Manufacture Date: SEP 21, 2023

Expiration Date: SEP 2026

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Sulfuric Acid	7664-93-9	ACS

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	Passed	
Assay (vs. Tris(hydroxymethyl)aminomethane/Methyl Red)	0.04995-0.05005 N at 20°C	0.05000 N at 20°C	723

Specification	Reference
Sulfuric acid, 0.05 N	TAPPI (T 610 sp-97)
Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.	

Part Number	Size / Package Type	Shelf Life (Unopened Container)
8219-1	4 L natural poly	36 months
8219-16	500 mL natural poly	36 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Sharon Travers (09/21/2023)

Operations Manager

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.