

Certificate of Analysis

Hydrogen Peroxide, 3% (w/w), Stabilized Reagent Grade

Lot Number: 2311H45

Product Number: 3819

Manufacture Date: NOV 15, 2023

Expiration Date: MAY 2025

Acetanilide is added as a preservative.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Hydrogen Peroxide	7722-84-1	ACS
Acetanilide	103-84-4	Reagent

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	Passed	
Assay (vs. Potassium Permanganate)	3.2-3.4 % (w/w)	3.4 % (w/w)	8040

Specification	Reference
Dilute Hydrogen Peroxide Solution	APHA (7500-Ra C)
Hydrogen Peroxide (1+9)	ASTM (D 3454)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
3819-1	4 L black poly	18 months
3819-16	500 mL amber poly	18 months
3819-32	1 L amber poly	18 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Sharon Travers (11/15/2023)

Operations Manager

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.