

Certificate of Analysis

Robert's Reagent, for the Qualitative Detection of Albumin in Urine

Lot Number: 2401F81

Product Number: 6630

Manufacture Date: JAN 16, 2024

Expiration Date: 2025-12-31

For In Vitro Diagnostic Use. This product is suitable for use in the qualitative ring or contact Robert's Test for Albumin in urine. Consult RICCA CHEMICAL COMPANY Technical Reference Document TRD16 for details.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Magnesium Sulfate Heptahydrate	10034-99-8	ACS
Nitric Acid	7697-37-2	ACS

Test	Specification	Result
Appearance	Colorless liquid	Passed
Refractive Index	To Pass Test	Passed

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
6630-16	500 mL amber glass	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

Jose Pena (01/16/2024)

Operations Manager

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.