

Certificate of Analysis

Sulfuric Acid, 0.100 Normal (N/10)

Lot Number: 2401J81

Product Number: 8250

Manufacture Date: JAN 29, 2024

Expiration Date: JAN 2027

The certified value for this product is confirmed in independent testing by a second qualified chemist.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Sulfuric Acid	7664-93-9	ACS

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	Passed	
Assay (vs. Tris(hydroxymethyl)aminomethane/Methyl Red)	0.0999-0.1001 N at 20°C	0.1000 N at 20°C	723

Specification	Reference
Sulfuric Acid, 0.05 M (0.1 N)	APHA (4500-SiO2 E)
Standard sulfuric acid, 0.1 N	APHA (2320 B)
Sulfuric Acid, 0.1 N	EPA (SW-846) (9315)
Standard Sulfuric Acid, 0.1 N	EPA (310.1)
Sulfuric Acid (0.1 N)	EPA (SW-846) (9057)
Sulfuric Acid (0.05 M)	EPA (SW-846) (50)
Sulfuric Acid (0.05 M)	EPA (SW-846) (51)
Sulfuric acid, 0.1 N	TAPPI (T 418 cm-97)
Sulfuric Acid, 0.1 N standard	EPA (SW-846) (9080)
Standard Sulfuric Acid, 0.1 N	EPA (SW-846) (9080)
Sulfuric Acid, Standard (0.05 M)	ASTM (D 3228)
Sulfuric acid, 0.1 N	TAPPI (T 610 sp-97)
Sulfuric Acid Solution (1+359)	ASTM (D 3454)
0.05 M Sulphuric acid Volumetric Solution	EP (3008000)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
8250-1	4 L natural poly	36 months
8250-16	500 mL natural poly	36 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Jose Pena (01/29/2024)

Operations Manager

This document is designed to comply with ISO Guide 31 "Reference Materials --
Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.