1841 Broad Street Pocomoke City, MD 21851 http://www.riccachemical.com 1-888-GO-RICCA

customerservice@riccachemical.com

Certificate of Analysis

Formalin-Aceto-Alcohol (F-A-A Mixture), for Rapid Fixation of Tissue

Lot Number: 2403F46 Product Number: 3210

Manufacture Date: MAR 07, 2024

Expiration Date: AUG 2025

For In Vitro Diagnostic Use. This product is suitable for rapid fixation of tissues (small pieces 2 mm in thickness are fixed in 4-6 hours), for preservation of glycogen and central nervous system tissues, and for cytoplasmic fixation when RNA digestion is to be preformed. This is not an idea fixative for tissue storage and routine use. This fixative prevents solution of carbohydrates before the protein component is completely fixed. The Acetic Acid in this fixative ensures fixation of nucleoprotein, providing an improved histological picture.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Ethyl Alcohol	64-17-5	Reagent
Formaldehyde	50-00-0	ACS
Methyl Alcohol	67-56-1	ACS, Reagent
Acetic Acid	64-19-7	ACS

Test	Specification	Result	
Appearance	Colorless liquid	Passed	

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
3210-16	500 mL natural poly	18 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

Jose Pena (03/07/2024)

Operations Manager

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

Version: 1.3 Lot Number: 2403F46 Product Number: 3210 Page 1 of 1