

# Certificate of Analysis

**Total Suspended Solids (TSS) Standard, 500 ppm TSS (Gravimetric, Dried at 103-105°C)**

**Lot Number:** 2404J36

**Product Number:** 8672.6

**Manufacture Date:** APR 24, 2024

**Expiration Date:** APR 2025

This product is suitable for use as a check standard for APHA Method 2540 D and EPA Method 160.2. The TSS value has a 95% confidence interval ( $\pm 2$  Standard Deviations) and is certified traceable to the NIST national mass standard. This standard will settle during storage. Invert bottle several times to mix well just before each use.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Cellulose	9004-34-6	Laboratory

Test	Specification	Result
Appearance	Colorless liquid	Passed
Total Suspended Solids (EPA 160.2)	465-535 mg/L	500 mg/L

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
8672.6-32	1 L amber glass	12 months

**Recommended Storage:** 15°C - 30°C (59°F - 86°F)



Jose Pena (04/24/2024)  
Operations Manager

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.