

Certificate of Analysis

Eriochrome Blue Black R Indicator, 0.2% (w/w) in Sodium Chloride

Lot Number: 4401S81

Product Number: 2902

Manufacture Date: JAN 29, 2024

Expiration Date: JAN 2026

Name	CAS#	Grade
Sodium Chloride	7647-14-5	ACS
Eriochrome Blue Black R	2538-85-4	Reagent

Test	Specification	Result
Appearance	Violet/purple solid	Passed
Suitability for Use	Purple (excess Ca or Mg) - Pure Blue (excess EDTA)	Passed

Specification	Reference
Eriochrome Blue Black R Indicator	APHA (3500-Ca B)
Eriochrome Blue Black R Indicator, dry powder	EPA (215.2)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
2902-100	100 g poly	24 months
2902-500	500 g poly natural	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

Paul Brandon (01/29/2024)

Production Manager

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.