1490 Lammers Pike Batesville, IN 47006 http://www.riccachemical.com 1-888-GO-RICCA

customerservice@riccachemical.com

Certificate of Analysis

ORP Standard, +200 mV

vs. Ag/AgCl (4M KCl sat. with AgCl filling solution)

Lot Number: 4402H81 Product Number: 5464.5

Manufacture Date: FEB 22, 2024

Expiration Date: NOV 2024

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Potassium Chloride	7447-40-7	ACS
Potassium Ferricyanide	13746-66-2	ACS
Potassium Ferrocyanide Trihydrate	14459-95-1	ACS

Test	Specification	Result
Appearance	Yellow-green liquid	Passed
Oxidation-Reduction Potential (ORP)	190-210 mV at 25°C	199 mV at 25°C

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
5464.5-16	500 mL amber glass	9 months
5464.5-2.5	10 L Cubitainer®	9 months
5464.5-4	120 mL amber glass	9 months

Recommended Storage: 2°C - 8°C (36°F - 46°F)

Paul Brandon (02/22/2024)

Production Manager

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

Version: 1.3 Lot Number: 4402H81 Product Number: 5464.5 Page 1 of 1