

# Certificate of Analysis

**EDTA Titrant, 0.100 Molar (M/10)**

**Lot Number:** 4403H48

**Product Number:** 2730

**Manufacture Date:** MAR 13, 2024

**Expiration Date:** MAR 2026

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Ethylenediaminetetraacetic Acid (EDTA), Disodium Salt, Dihydrate	6381-92-6	ACS

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	Passed	
Assay (vs. Calcium Carbonate/Eriochrome Black T)	0.0999-0.1001 Molar	0.1000 Molar	915

Specification	Reference
---------------	-----------

EDTA, 0.1 Molar Volumetric Solution

ACS (N/A)

EDTA Solution (0.1 N)

ASTM (D 2036 X2)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
2730-2.5	10 L Cubitainer®	24 months
2730-32	1 L natural poly	24 months
2730-5	20 L Cubitainer®	24 months

**Recommended Storage:** 15°C - 30°C (59°F - 86°F)



Paul Brandon (03/13/2024)

Production Manager

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.