

Certificate of Analysis

Mercuric Nitrate, 0.0100 Normal (N/100)

Lot Number: 4403H93	Product Number: 4704	Manufacture Date: MAR 14, 2024 Expiration Date: MAR 2026	
Name	CAS#	Grade	
Water	7732-18-5	ACS/ASTM/USP/EP	
Nitric Acid	7697-37-2	ACS	
Mercuric Nitrate Monohydrate	7783-34-8	ACS	
T+		D14	NIGT SDM#

Test	Specification	Result	NISI SRM#
Appearance	Colorless liquid	Passed	
Assay (vs. Potassium Chloride/Diphenylcarbazone)	0.00999-0.01001 N at 20°C	0.00000 N at 20°C	999

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
4704-1	4 L amber glass	24 months
4704-32	1 L amber glass	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

Hand Dranda

Paul Brandon (03/14/2024) Production Manager This document is designed to comply with ISO Guide 31 "Reference Materials --Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.