

Product Specification

Acetonitrile with 0.01% Formic Acid (v/v)

Lot Number: SAMPLE

Product Number: R0234100

Manufacture Date: N/A

Expiration Date: N/A

Name	CAS#	Grade
Formic Acid	64-18-6	HPLC
Water	7732-18-5	HPLC
Acetonitrile	75-05-8	HPLC

Test	Specification	Result
Appearance	Colorless liquid	N/A

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
R0234100-4C	4 L amber glass	12 months
R0234100-4x4C	4 x 4 L amber glass	12 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.