

Product Specification

Ceric Sulfate, 0.4 Normal in 2 Normal Sulfuric Acid

Lot Number: SAMPLE

Product Number: R1910000

Manufacture Date: N/A

Expiration Date: N/A

Name	CAS#	Grade
Water	7732-18-5	Reagent
Sulfuric Acid	7664-93-9	Reagent
Ceric Sulfate	13590-82-4	Reagent

Test	Specification	Result
Appearance	Orange liquid	N/A
Assay	Actual Value Reported	N/A
Identification	To Pass Test	N/A

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
R1910000-20F	20 L Cubitainer [®]	36 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.