

# Product Specification

## Ethanol, 95.5% USP

**Lot Number:** SAMPLE

**Product Number:** R2916000

**Date Packaged:** N/A

**Expiration Date:** N/A

|                 |
|-----------------|
| <b>Contains</b> |
|-----------------|

 $C_2H_5OH$  CAS No. 64-17-5 MW 46.07

| Test                     | Specification          | Result |
|--------------------------|------------------------|--------|
| Acidity or Alkalinity    | To Pass Test           | N/A    |
| Appearance               | Colorless liquid       | N/A    |
| Assay                    | 94.9-96.0 % (v/v)      | N/A    |
| Clarity                  | To Pass Test           | N/A    |
| Color                    | To Pass Test           | N/A    |
| Identification - Ethanol | To Pass Test           | N/A    |
| Nonvolatile Residue      | To Pass Test           | N/A    |
| Specific Gravity         | 0.812-0.816 at 15.56°C | N/A    |
| UV Absorption            | To Pass Test           | N/A    |
| Volatile Impurities      | To Pass Test           | N/A    |

| Specification | Reference |
|---------------|-----------|
|---------------|-----------|

|         |               |
|---------|---------------|
| Alcohol | USP (Current) |
|---------|---------------|

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

| Part Number   | Size / Package Type | Shelf Life (Unopened Container) |
|---------------|---------------------|---------------------------------|
| R2916000-1A   | 1 L natural poly    | 36 months                       |
| R2916000-4A   | 4 L natural poly    | 36 months                       |
| R2916000-4C   | 4 L amber glass     | 36 months                       |
| R2916000-500A | 500 mL natural poly | 36 months                       |
| R2916000-55E  | 55 gal poly drum    | 36 months                       |

**Recommended Storage:** 15°C - 30°C (59°F - 86°F)

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.