

# Product Specification

## Nitrification Inhibitor

Lot Number: SAMPLE

Product Number: R5443990

Manufacture Date: N/A

Expiration Date: N/A

Name	CAS#	Grade
Sodium Sulfate Anhydrous	7757-82-6	ACS
2-Chloro-6-(trichloromethyl)pyridine	1929-82-4	Purified

Test	Specification	Result
------	---------------	--------

Appearance	Off-white solid	N/A
------------	-----------------	-----

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
R5443990-125D2	125 g glass amber	24 months
R5443990-500C2	500 g amber glass	24 months
R5443990-50B1	50 g poly	24 months

**Recommended Storage:** 15°C - 30°C (59°F - 86°F)

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.