

Product Specification

Sudan III, Saturated Alcoholic Solution Fat Stain for Microscopic Examination of Feces

Lot Number: SAMPLE Product Number: 8086

Manufacture Date: N/A

Expiration Date: N/A

For In Vitro Diagnostic Use. May be suitable for other laboratory applications.

Name	CAS#	Grade
Ethyl Alcohol	64-17-5	Reagent
Methyl Alcohol	67-56-1	Reagent
Sudan III	85-86-9	Certified

Test	Specification	Result
Appearance	Red liquid	N/A

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
8086-16	500 mL natural poly	12 months
8086-32	1 L natural poly	12 months
8086-4	120 mL natural poly	12 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.