Arlington, TX 76012
Pocomoke City, MD 21851
Batesville, IN 47006
http://www.riccachemical.com
1-888-GO-RICCA
customerservice@riccachemical.com

Product Specification

Sulfuric Acid, 0.500 Normal (N/2)

Lot Number: SAMPLE Product Number: 8280

Manufacture Date: N/A

Expiration Date: N/A

| Name | CAS# | Grade | |
|---------------|-----------|-----------------|--|
| Water | 7732-18-5 | ACS/ASTM/USP/EP | |
| Sulfuric Acid | 7664-93-9 | ACS | |

| Test | Specification | Result | NIST SRM# |
|---|-------------------------|--------|-----------|
| Appearance | Colorless liquid | N/A | |
| Assay (vs. | 0.4998-0.5002 N at 20°C | N/A | 723 |
| Tris(hydroxymethyl)aminomethane/Methyl Red) | | | |

| Specification | Reference |
|---------------------------------------|----------------------|
| Sulfuric acid, 0.5 N | TAPPI (T 624 cm-85) |
| Standard sulfuric acid, 0.5 N | TAPPI (T 621 cm-71) |
| Sulfuric acid solution, 0.5 N | TAPPI (T 600 cm-83) |
| Sulfuric Acid, 0.5 N | EPA (SW-846) (7470A) |
| Sulfuric Acid, 0.5 N | EPA (SW-846) (7471A) |
| Sulfuric Acid, 0.5 N | EPA (245.1) |
| Sulfuric Acid, 0.5 N | EPA (245.5) |
| Sulfuric acid, 0.5 N | TAPPI (T 610 sp-97) |
| Sulfuric Acid, 0.5 N Reagent Solution | ACS (N/A) |

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

| Part Number | Size / Package Type | Shelf Life (Unopened Container) |
|-------------|---------------------|---------------------------------|
| 8280-1 | 4 L natural poly | 36 months |
| 8280-16 | 500 mL natural poly | 36 months |
| 8280-1CT | 4 L Cubitainer® | 36 months |
| 8280-2.5 | 10 L Cubitainer® | 36 months |
| 8280-32 | 1 L natural poly | 36 months |
| 8280-4 | 120 mL natural poly | 36 months |
| 8280-5 | 20 L Cubitainer® | 36 months |
| 8280-55 | 55 gal poly drum | 36 months |

Recommended Storage: 15°C - 30°C (59°F - 86°F)

Version: 1.3 Lot Number: SAMPLE Product Number: 8280 Page 1 of 2

