

# Product Specification

**Total Suspended Solids (TSS) Standard, 200 ppm TSS (Gravimetric, Dried at 103-105°C)****Lot Number:** SAMPLE**Product Number:** 8672.5**Manufacture Date:** N/A**Expiration Date:** N/A

This product is suitable for use as a check standard for APHA Method 2540 D and EPA Method 160.2. The TSS value has a 95% confidence interval ( $\pm 2$  Standard Deviations) and is certified traceable to the NIST national mass standard.  
This standard will settle during storage. Invert bottle several times to mix well just before each use.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Cellulose	9004-34-6	Laboratory

Test	Specification	Result
Appearance	Colorless liquid	N/A
Total Suspended Solids (EPA 160.2)	186-214 mg/L	N/A

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
8672.5-16	500 mL amber glass	12 months
8672.5-32	1 L amber glass	12 months

**Recommended Storage:** 15°C - 30°C (59°F - 86°F)

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.