

# Product Specification

## HydroSpec<sup>®</sup> Solvent K

### Medium for One and Two Component Volumetric Titrations of Aldehydes and Ketones

Lot Number: SAMPLE

Product Number: RK340000

Manufacture Date: N/A

Recommended Retest Date: N/A

| Name                              | CAS#      | Grade |
|-----------------------------------|-----------|-------|
| Chloroform                        | 67-66-3   |       |
| Sulfur Dioxide                    | 7446-09-5 |       |
| Propylene Glycol Monomethyl Ether | 107-98-2  |       |

| Test                     | Specification          | Result |
|--------------------------|------------------------|--------|
| Suitability for Use      | Karl Fischer titration | N/A    |
| Water (H <sub>2</sub> O) | 0.002-0.030            | N/A    |

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

| Part Number | Size / Package Type | Shelf Life (Unopened Container) |
|-------------|---------------------|---------------------------------|
| RK340000-1C | 1 L amber glass     | 60 months                       |

**Recommended Storage:** 15°C - 30°C (59°F - 86°F)

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.