448 West Fork Dr Arlington, TX 76012 http://www.riccachemical.com 1-888-GO-RICCA

customerservice@riccachemical.com

## Certificate of Analysis

Isopropyl Alcohol, 70% (v/v)

Lot Number: 1311F18 Product Number: 4210

Manufacture Date: NOV 14, 2023

Expiration Date: OCT 2025

This product is intended for laboratory cleaning and disinfection purposes and is typically used for disinfection of surfaces. All components of this product are filtered as part of the manufacturing process. The residue after evaporation is typically 0.001% (10 ppm) maximum at the time of manufacturing. A culture/colony count of 1 CFU/mL, or less, is typical for the water at the time of manufacturing.

Name	CAS#	Grade	
Isopropyl Alcohol	67-63-0	ACS/USP	
Water	7732-18-5	ACS/ASTM/USP/EP	

Test	Specification	Result
Appearance	Colorless liquid	Passed
Assay (by Specific Gravity)	68-72 % (v/v)	69 % ( <sub>v/v</sub> )
Specific Gravity at 20°C	0.874-0.882	0.881

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
4210-1	4 L natural poly	24 months
4210-16	500 mL natural poly	24 months
4210-32	1 L natural poly	24 months
4210-5	20 L Cubitainer®	24 months
4210-5HP	$20~\mathrm{L~Ropak^{TM}}$	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

Heidi J Green (11/14/2023)

**Operations Manager** 

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

Version: 1.5 Lot Number: 1311F18 Product Number: 4210 Page 1 of 1