

Certificate of Analysis

Potassium Hydroxide, 0.100 Normal (N/10)

Lot Number: 1401H97

Product Number: 6200

Manufacture Date: JAN 23, 2024

Expiration Date: JAN 2026

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Potassium Hydroxide	1310-58-3	ACS

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	Passed	
Assay (vs. Potassium Hydrogen Phthalate/Phenolphthalein)	0.0999-0.1001 N at 20°C	0.0999 N at 20°C	84

Specification	Reference
Potassium Hydroxide, 0.1 M	EPA (SW-846) (61)
Potassium Hydroxide, 0.1 N aqueous solution	ASTM (D 2619)
Potassium hydroxide, 0.1 N	TAPPI (T 610 sp-97)
Potassium hydroxide, 0.1 N	TAPPI (T 689 om-93)
0.1 M Potassium hydroxide Volumetric Solution	EP (3004800)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
6200-1	4 L natural poly	24 months
6200-16	500 mL natural poly	24 months
6200-32	1 L natural poly	24 months
6200-5	20 L Cubitainer®	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Heidi J Green (01/23/2024)

Operations Manager

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.