448 West Fork Dr Arlington, TX 76012 http://www.riccachemical.com 1-888-GO-RICCA

customerservice@riccachemical.com

Certificate of Analysis

Gentian Violet, 1% (w/v)

Lot Number: 1402989 Product Number: 3230

Manufacture Date: FEB 08, 2024

Expiration Date: JAN 2026

For In Vitro Diagnostic Use. May be suitable for other laboratory applications.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Gentian Violet	548-62-9	Certified

Test	Specification	Result	
Appearance	Dark violet liquid	Passed	
UV/Visible Spectrum	To Pass Test	Passed	

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
3230-16	500 mL natural poly	24 months
3230-4	120 mL natural poly	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

Heidi J Green (02/08/2024)

Operations Manager

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

Version: 1.3 Lot Number: 1402989 Product Number: 3230 Page 1 of 1