RICCA CHEMICAL COMPANY

Lot Number: 1402C05

Product Number: 7996

Manufacture Date: FEB 14, 2024

Expiration Date: AUG 2025

For In Vitro Diagnostic Use. This product is intended for dilution and staining of cerebrospinal fluid samples for moderate total cell counts by microscopic examination.

Recommended procedure: The cells should be counted as soon as the specimen is received. Mix the specimen thoroughly. Draw the spinal fluid to mark 1 in a white cell counting pipet and add this diluting fluid to the 11 mark (1:10 dilution). Mix, discard one-third, and place 1 drop on each side of the blood counting chamber. Count the cells in five squares for each drop (four corners and center), and add the total for all 10 squares counted. Multiply by 10 to obtain the count in cells/mm³. This diluting fluid is not suitable for use with very cloudy or bloody samples.

Name	CAS#	Grade	
Water	7732-18-5	ACS/ASTM/USP/EP	
Acetic Acid	64-19-7	ACS	
Gentian Violet	548-62-9	Certified	

Test	Specification	Result
Appearance	Dark violet liquid	Passed

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
7996-16	500 mL natural poly	18 months
7996-32	1 L natural poly	18 months
7996-4	120 mL natural poly	18 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

Heidi J Green (02/14/2024) Operations Manager

This document is designed to comply with ISO Guide 31 "Reference Materials --Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.