448 West Fork Dr Arlington, TX 76012 http://www.riccachemical.com 1-888-GO-RICCA

customerservice@riccachemical.com

Certificate of Analysis

Papanicolaou Stain, OG-6 Formula, Modified Counterstain for use with Hematoxylin Stains in Papanicolaou Staining

Lot Number: 1403E65 Product Number: 5510

Manufacture Date: MAR 06, 2024

Expiration Date: FEB 2026

For In Vitro Diagnostic Use. This product is suitable for use in various regressive and progressive Papanicolaou staining procedures. Consult RICCA CHEMICAL COMPANY Technical Reference Document TRD07 for a typical staining procedure. Other staining procedures using this reagent may be found in standard clinical laboratory texts and operating procedures. Filtration may be required before use.

Name	CAS#	Grade
Ethyl Alcohol	64-17-5	Reagent
Water	7732-18-5	ACS/ASTM/USP/EP
Methyl Alcohol	67-56-1	Reagent
Phosphotungstic Acid Hydrate	12501-23-4	Reagent
Orange G	1936-15-8	Certified

Test	Specification	Result	
Appearance	Orange liquid	Passed	_
UV/Visible Spectrum	To Pass Test	Passed	

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
5510-1	4 L natural poly	24 months
5510-16	500 mL natural poly	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

Heidi J Green (03/06/2024)

Operations Manager

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

Version: 1.3 Lot Number: 1403E65 Product Number: 5510 Page 1 of 1