

Certificate of Analysis

Gardner Color Standard #1

Lot Number: 1403J03 Pro	duct Number: 3226.01	Manufacture Date: MAR 19, 2024 Expiration Date: MAR 2025
Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Hydrochloric Acid	7647-01-0	ACS
Potassium Hexachloroplatinate (IV)	16921-30-5	Reagent
Test	Specification	Result
Appearance	Yellow liquid	Passed
Gardner color (ASTM D 6166)	0.7-1.3	1.0
Specification	Reference	

Specification

Gardner Color Standard No. 1

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

ASTM (D 1544-58 T)

Part Number	Size / Package Type	Shelf Life (Unopened Container)		
3226.01-16	500 mL amber glass	12 months		
D ecommonded Storege: $15^{\circ}C = 20^{\circ}C$ ($50^{\circ}E = 86^{\circ}E$)				

Recommended Storage: 15°C - 30°C (59°F - 86°F)

Heidi J Green (03/19/2024) **Operations Manager** This document is designed to comply with ISO Guide 31 "Reference Materials --Contents of Certificates and Labels.

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.