448 West Fork Dr Arlington, TX 76012 http://www.riccachemical.com 1-888-GO-RICCA

customerservice@riccachemical.com

## Certificate of Analysis

Iodine (Iodine-Iodide), 0.100 Normal (N/10)

Lot Number: 1404F54 Product Number: 4000

Manufacture Date: APR 09, 2024

Expiration Date: JAN 2025

The certified value for this product is confirmed in independent testing by a second qualified chemist.

To meet USP specifications for Iodine, Tenth-Normal (0.1 N) Volumetric Solution (VS) and Test Solution (TS) add 3-drops of Hydrochloric acid per Liter of R4000.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Potassium Iodide	7681-11-0	ACS
Iodine	7553-56-2	ACS

Test	Specification	Result	NIST SRM#
Appearance	Dark brown liquid	Passed	
Assay (vs. Sodium Thiosulfate/Starch)	0.0998-0.1002 N at 20°C	0.1001 N at 20°C	136

Specification	Reference
Standard Iodine, 0.1 N	APHA (4500-Cl B)
Standard iodine Solution, 0.1 N	APHA (4500-C1 C)
Iodine Solution	APHA (3500-Pb B)
Standard Iodine Solution (0.1 N)	EPA (330.2)
Iodine, 0.1 N	TAPPI (T 622 cm-84)
Standard Iodine Solution (0.1 N)	EPA (330.3)
Iodine, 0.1 Normal Volumetric Solution	ACS (N/A)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
4000-4	120 mL amber glass	12 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

Version: 1.3 Lot Number: 1404F54 Product Number: 4000 Page 1 of 2

Heidi J Green (04/09/2024)

Operations Manager

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

Version: 1.3 Lot Number: 1404F54 Product Number: 4000 Page 2 of 2