

Certificate of Analysis

Potassium Permanganate, 0.0100 Normal (N/100)

Lot Number: 1404G91

Product Number: 6370

Manufacture Date: APR 15, 2024 Expiration Date: OCT 2025

The certified value reported is the prepared value based upon the method of preparation of the material. The uncertainty in the prepared value is based upon the volumetric method of preparation.

Name	CAS#	Grade	
Water	7732-18-5	ACS/ASTM/USP/EP	
Potassium Permanganate	7722-64-7	ACS	

Test	Specification	Result	NIST SRM#
Appearance	Purple liquid	Passed	
Potassium Permanganate	0.0099-0.0101 N at 20°C	0.0100 N at 20°C	8040

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
6370-16	500 mL amber glass	18 months
6370-32	1 L amber glass	18 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

Heidi J Green (04/15/2024) Operations Manager This document is designed to comply with ISO Guide 31 "Reference Materials --Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.