

Certificate of Analysis

Sulfuric Acid, 0.0200 Normal (N/50)

Lot Number: 1404H52

Product Number: 8200

Manufacture Date: APR 16, 2024

Expiration Date: MAR 2027

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Sulfuric Acid	7664-93-9	ACS

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	Passed	
Assay (vs. Tris(hydroxymethyl)aminomethane/Methyl Red)	0.01998-0.02002 N at 20°C	0.02001 N at 20°C	723

Specification	Reference
Sulfuric Acid, Standard (0.02 N)	ASTM (D 1067 A)
Sulfuric Acid Solution, Standard (0.02 N, 1 mL = 0.28 mg ammonia nitrogen)	ASTM (D 3590 A)
Sulfuric Acid, Standard (0.02 N)	ASTM (D 1067 B)
Sulfuric Acid, Standard (0.02 N)	ASTM (D 1067 C)
Standard sulfuric acid, 0.02 N	APHA (2320 B)
Sulfuric Acid, 0.02 N	APHA (3500-A1 B)
Standard Sulfuric Acid Titrant, 0.02 N	APHA (4500-NH3 C)
Standard Sulfuric Acid, 0.02 N	EPA (310.1)
Dilute Sulfuric Acid (0.01 M)	EPA (SW-846) (9212)
Dilute Sulfuric Acid (0.01 M)	EPA (SW-846) (9211)
Standard Sulfuric Acid, 0.02 N	EPA (305.1)
Sulfuric Acid, Standard Solution (0.02 N)	EPA (350.2)
Sulfuric Acid, Standard Solution (0.02 N)	EPA (351.3)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
8200-2.5	10 L Cubitainer®	36 months
8200-5	20 L Cubitainer®	36 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Heidi J Green (04/16/2024)
Operations Manager

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.