

Certificate of Analysis

Sulfuric Acid, 0.500 Normal (N/2)

Lot Number: 1404L75

Product Number: 8280

Manufacture Date: APR 10, 2024

Expiration Date: APR 2027

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Sulfuric Acid	7664-93-9	ACS

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	Passed	
Assay (vs. Tris(hydroxymethyl)aminomethane/Methyl Red)	0.4998-0.5002 N at 20°C	0.5000 N at 20°C	723

Specification	Reference
Sulfuric acid, 0.5 N	TAPPI (T 624 cm-85)
Standard sulfuric acid, 0.5 N	TAPPI (T 621 cm-71)
Sulfuric acid solution, 0.5 N	TAPPI (T 600 cm-83)
Sulfuric Acid, 0.5 N	EPA (SW-846) (7470A)
Sulfuric Acid, 0.5 N	EPA (SW-846) (7471A)
Sulfuric Acid, 0.5 N	EPA (245.1)
Sulfuric Acid, 0.5 N	EPA (245.5)
Sulfuric acid, 0.5 N	TAPPI (T 610 sp-97)
Sulfuric Acid, 0.5 N Reagent Solution	ACS (N/A)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
8280-1	4 L natural poly	36 months
8280-32	1 L natural poly	36 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Heidi J Green (04/10/2024)
Operations Manager

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.