

Certificate of Analysis

Chloride Standard, 100,000 ppm Cl⁻ (2.82 Normal)

Lot Number: 1405C35

Product Number: 1956

Manufacture Date: MAY 01, 2024

Expiration Date: APR 2026

| Name | CAS# | Grade |
|-----------------|-----------|-----------------|
| Water | 7732-18-5 | ACS/ASTM/USP/EP |
| Sodium Chloride | 7647-14-5 | ACS |

| Test | Specification | Result | NIST SRM# |
|---|---------------------|---------------|-----------|
| Appearance | Colorless liquid | Passed | |
| Assay (vs. Silver Nitrate/Potassium Chromate) | 99900-100100 ppm Cl | 100084 ppm Cl | 999 |

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

| Part Number | Size / Package Type | Shelf Life (Unopened Container) |
|-------------|---------------------|---------------------------------|
| 1956-16 | 500 mL natural poly | 24 months |
| 1956-32 | 1 L natural poly | 24 months |

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Heidi J Green (05/01/2024)

Operations Manager

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.