

Certificate of Analysis

Potassium Hydroxide, 0.500 Normal (N/2) in 90% (v/v) Denatured Alcohol

Lot Number: 2404J33

Product Number: 6263

Manufacture Date: APR 24, 2024

Expiration Date: APR 2025

Name	CAS#	Grade
Ethyl Alcohol	64-17-5	Reagent
Water	7732-18-5	ACS/ASTM/USP/EP
Methyl Alcohol	67-56-1	Reagent
Potassium Hydroxide	1310-58-3	Reagent
Preservative	Proprietary	

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	Passed	
Assay (vs. Sulfuric Acid/Phenolphthalein)	0.4995-0.5005 N at 20°C	0.5000 N at 20°C	723

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
6263-32	1 L natural poly	12 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)Jose Pena (04/24/2024)
Operations Manager

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.