

Certificate of Analysis

Water Hardness Indicator, 0.5% (w/v) Eriochrome Black T in 2-Methoxyethanol

Lot Number: 2407F68

Product Number: 9300

Manufacture Date: JUL 16, 2024

Expiration Date: APR 2025

Name	CAS#	Grade
Ethylene Glycol Monomethyl Ether	109-86-4	ACS
Eriochrome Black T	1787-61-7	ACS

Test	Specification	Result
Appearance	Dark purple liquid	Passed
Suitability for Use	Wine Red (excess Ca or Mg) - Pure Blue (excess EDTA)	Passed

Specification	Reference
Eriochrome Black T Indicator	APHA (2340 C) EPA (215.2)
Eriochrome Black T Indicator	EPA (130.2)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
9300-16	500 mL amber glass	9 months
9300-32	1 L amber glass	9 months
9300-4	120 mL amber glass	9 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)Jose Pena (07/16/2024)
Operations Manager

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.