

Certificate of Analysis

Citric Acid, 1.00 Molar

Lot Number: 4401K03

Product Number: 2131

Manufacture Date: JAN 15, 2024

Expiration Date: JAN 2025

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Citric Acid	77-92-9	USP

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	Passed	
Assay (vs. Sodium Hydroxide/Phenolphthalein)	0.999-1.001 Molar	1.001 Molar	84

Specification	Reference
Citric Acid, 1 M	APHA (7500-Ra B)
Citric Acid, 1 M	APHA (7500-Ra D)
Citric Acid, 1 M	EPA (SW-846) (9315)
Citric Acid, 1.0 M	EPA (SW-846) (8315A)
Citric Acid, 1 M	EPA (SW-846) (9320)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
2131-1	4 L natural poly	12 months
2131-16	500 mL natural poly	12 months
2131-32	1 L natural poly	12 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Paul Brandon (01/15/2024)

Production Manager

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.